

Result: Analysis Table

ID: Zeolite F-fine 19/6	Run No: 6	Measured: 22/6/2015 3:52ii
File: 200969	Rec. No: 6	Analysed: 22/6/2015 3:52ii
Path: C:\SIZERS\DATA\CLIENTS\2015\06\OLYMPOS\		Source: Analysed

Range: 300RF mm	Beam: 2.40 mm	Sampler: MS14	Obs': 23.9 %
Presentation: 3OHD		Analysis: Polydisperse	Residual: 0.369 %
Modifications: None			

Conc. = 0.0119 %Vol	Density = 1.000 g/cm ³	S.S.A. = 2.4741 m ² /g
Distribution: Volume	D[4, 3] = 8.90 um	D[3, 2] = 2.43 um
D(v, 0.1) = 1.12 um	D(v, 0.5) = 5.79 um	D(v, 0.9) = 20.76 um
Span = 3.392E+00	Uniformity = 1.081E+00	

Size (um)	Volume Under%	Size (um)	Volume Under%	Size (um)	Volume Under%	Size (um)	Volume Under%
0.05	0.00	0.67	5.33	9.00	65.10	120.67	100.00
0.06	0.00	0.78	6.37	10.48	70.34	140.58	100.00
0.07	0.00	0.91	7.71	12.21	75.41	163.77	100.00
0.08	0.01	1.06	9.35	14.22	80.19	190.80	100.00
0.09	0.02	1.24	11.36	16.57	84.55	222.28	100.00
0.11	0.04	1.44	13.76	19.31	88.39	258.95	100.00
0.13	0.08	1.68	16.54	22.49	91.62	301.68	100.00
0.15	0.15	1.95	19.64	26.20	94.22	351.46	100.00
0.17	0.27	2.28	23.08	30.53	96.21	409.45	100.00
0.20	0.47	2.65	26.82	35.56	97.65	477.01	100.00
0.23	0.77	3.09	30.85	41.43	98.65	555.71	100.00
0.27	1.20	3.60	35.14	48.27	99.30	647.41	100.00
0.31	1.74	4.19	39.67	56.23	99.70	754.23	100.00
0.36	2.35	4.88	44.44	65.51	99.93	878.67	100.00
0.42	2.99	5.69	49.41	76.32	100.00		
0.49	3.68	6.63	54.55	88.91	100.00		
0.58	4.45	7.72	59.80	103.58	100.00		

