

Result: Analysis Table

ID: Diat+Zeo Fine	Run No: 18	Measured: 16/9/2015 11:46:01
File: 204298A	Rec. No: 18	Analysed: 16/9/2015 11:46:01
Path: C:\SIZERS\DATA\CLIENTS\2015\09\OLYMPOS\		Source: Analysed

Range: 300RF mm	Beam: 2.40 mm	Sampler: MS14	Obs': 11.6 %
Presentation: 3OHD		Analysis: Polydisperse	Residual: 0.364 %
Modifications: None			

Conc. = 0.0056 %Vol	Density = 1.000 g/cm ³	S.S.A.= 2.4098 m ² /g
Distribution: Volume	D[4, 3] = 8.76 um	D[3, 2] = 2.49 um
D(v, 0.1) = 1.21 um	D(v, 0.5) = 5.72 um	D(v, 0.9) = 20.35 um
Span = 3.347E+00	Uniformity = 1.052E+00	

Size (um)	Volume Under%	Size (um)	Volume Under%	Size (um)	Volume Under%	Size (um)	Volume Under%
0.05	0.00	0.67	5.25	9.00	66.65	120.67	100.00
0.06	0.00	0.78	6.14	10.48	71.92	140.58	100.00
0.07	0.00	0.91	7.25	12.21	76.85	163.77	100.00
0.08	0.00	1.06	8.62	14.22	81.36	190.80	100.00
0.09	0.01	1.24	10.30	16.57	85.39	222.28	100.00
0.11	0.03	1.44	12.36	19.31	88.91	258.95	100.00
0.13	0.06	1.68	14.84	22.49	91.90	301.68	100.00
0.15	0.13	1.95	17.73	26.20	94.35	351.46	100.00
0.17	0.26	2.28	21.07	30.53	96.28	409.45	100.00
0.20	0.47	2.65	24.89	35.56	97.72	477.01	100.00
0.23	0.79	3.09	29.15	41.43	98.72	555.71	100.00
0.27	1.23	3.60	33.83	48.27	99.36	647.41	100.00
0.31	1.79	4.19	38.89	56.23	99.73	754.23	100.00
0.36	2.41	4.88	44.25	65.51	99.91	878.67	100.00
0.42	3.06	5.69	49.81	76.32	100.00		
0.49	3.74	6.63	55.48	88.91	100.00		
0.58	4.47	7.72	61.13	103.58	100.00		

